

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/05.10/6.1

Technical Data Sheet		Rosenberger
SMP	Calibration Kit Industrial Version	19CK010-150



Contents

Device	Part number	Quantity	Calibration Option ^a
Open circuit plug	19S12L-000D3	1	FC
Open circuit jack	19K12L-000D3	1	FC
Short circuit plug	19S12S-000D3	1	FC
Short circuit jack	19K12S-000D3	1	FC
Calibration load plug	19S150-C10D3	1	FC
Calibration load jack	19K150-C10D3	1	FC
Calibration adaptor plug/plug	19S121-S20D3	1	FC
Calibration adaptor jack/jack	19K121-K20D3	1	FC
Calibration adaptor plug/jack	19S121-K20D3	1	FC
Calibration adaptor RPC-2.92 plug to SMP plug	02S119-S20D3	1	FC
Calibration adaptor RPC-2.92 jack to SMP jack	02K119-K20D3	1	FC

a. See "Declaration of calibration options" for explanation.

Documentation

This kit is delivered with

- **USB-Stick**
Standard Definitions as data files for Vector Network Analyzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz). Calibration Certificate as PDF-file.
- **Standard Definitions Cards**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers.
- **Kit Info Card**
Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.
- **Calibration Certificate**
Details see "Declaration of calibration options"
- **Operating Manual**

Electrical specifications

This specification covers electrical key specifications for the calibration standards of the calibration kit. Specific and more detailed datasheets are available for each kit device among the part number.

Calibration standard	Frequency	Parameter	Specification
Opens^b (plug and jack)	DC to ≤ 4 GHz > 4 to ≤ 18 GHz > 18 to ≤ 40 GHz	Error from Nominal Phase	≤ 3.0° ≤ 5.0° ≤ 8.0°
Shorts^b (plug and jack)	DC to ≤ 4 GHz > 4 to ≤ 18 GHz > 18 to ≤ 40 GHz	Error from Nominal Phase	≤ 2.5° ≤ 6.0° ≤ 8.0°
Calibration loads (plug and jack)	DC to ≤ 4 GHz > 4 to ≤ 40 GHz	Return Loss	≥ 35 dB ≥ 25 dB
Calibration adaptors (plug/plug , jack/jack, plug/jack)	DC to ≤ 4 GHz > 4 to ≤ 18 GHz > 18 to ≤ 40 GHz	Return Loss	≥ 26 dB ≥ 20 dB ≥ 18 dB
Calibration adaptors RPC-2.92 to SMP (plug/plug and jack/jack)	DC to ≤ 12 GHz > 12 to ≤ 26.5 GHz > 26.5 to ≤ 40 GHz	Return Loss	≥ 32 dB ≥ 26 dB ≥ 21 dB

- b. The specifications for opens and shorts are given as allowed deviation from nominal model as defined in calibration certificate included with your kit.

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/05.10/6.1

Technical Data Sheet				Rosenberger			
SMP		Calibration Kit Industrial Version		19CK010-150			
Declaration of calibration options Factory Calibration Standard delivery for this kit includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, traceable to Rosenberger standards, national / international standards are not available. Model based standard definitions of the calibration standards are reported in Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format. Accredited Calibration Not available. For further, more detailed information see application note AN001 on the Rosenberger homepage. Calibration interval Recommendation12 months Recommended accessories - Rosenberger Test Port Adaptor - Rosenberger VNA Test cable kit and Microwave Cable Assemblies For further, more detailed information please visit our homepage www.rosenberger.com.							
While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.							
Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
M. Scherbauer	23.11.15	J_Krautenbacher	12.07.16	c00	15-1629	I_Wallner	12.07.16
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Email : info@rosenberger.de		Page 3 / 3

Компания «Life Electronics» занимается поставками электронных компонентов импортного и отечественного производства от производителей и со складов крупных дистрибьюторов Европы, Америки и Азии.

С конца 2013 года компания активно расширяет линейку поставок компонентов по направлению коаксиальный кабель, кварцевые генераторы и конденсаторы (керамические, пленочные, электролитические), за счёт заключения дистрибьюторских договоров

Мы предлагаем:

- Конкурентоспособные цены и скидки постоянным клиентам.
- Специальные условия для постоянных клиентов.
- Подбор аналогов.
- Поставку компонентов в любых объемах, удовлетворяющих вашим потребностям.
- Приемлемые сроки поставки, возможна ускоренная поставка.
- Доставку товара в любую точку России и стран СНГ.
- Комплексную поставку.
- Работу по проектам и поставку образцов.
- Формирование склада под заказчика.
- Сертификаты соответствия на поставляемую продукцию (по желанию клиента).
- Тестирование поставляемой продукции.
- Поставку компонентов, требующих военную и космическую приемку.
- Входной контроль качества.
- Наличие сертификата ISO.

В составе нашей компании организован Конструкторский отдел, призванный помогать разработчикам, и инженерам.

Конструкторский отдел помогает осуществить:

- Регистрацию проекта у производителя компонентов.
- Техническую поддержку проекта.
- Защиту от снятия компонента с производства.
- Оценку стоимости проекта по компонентам.
- Изготовление тестовой платы монтаж и пусконаладочные работы.



Тел: +7 (812) 336 43 04 (многоканальный)

Email: org@lifeelectronics.ru

www.lifeelectronics.ru